

Torque Certification

CERTIFICATE: 58344
 DATE: 1-May-15
 CUSTOMER: K&S Power Tongs
 MODEL: 6" Tong Torque Assembly
 MANUFACTURER: CDS
 SERIAL NO: 1131

	SHOULD BE Ft.Lbs.	ACTUAL Ft.Lbs.
0	0	0
3000	3000	3000
6000	6000	6000
9000	9000	9000
12000	12000	12000

TECHNICIAN: Q.K. Q.K.

COMMENTS: Calibrated for a 25" handle & 4.4301 compression pad

Test Gauge s/n JW19971106-1 Cal. Date Jan. 21, 2015

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1

