

## **Torque Certification**

CERTIFICATE:	40013
DATE:	12-Dec-12
CUSTOMER:	K&S Power Tongs
MODEL: MANUFACTURER: SERIAL NO:	6" Gauge Tong Torque Gauge M18810
	SHOULD BE Ft. Lbs 0 3000 6000 9000 12000

TECHNICIAN: OK

COMMENTS: Calibrated for a 25" handle & 4.4301 compression pad

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1



5921 - 87A STREET, EDMONTON, ALBERTA T6E 5W6 TELEPHONE (780) 465-6006 FAX (780) 469-8872