

## Torque Certification

CERTIFICATE: WO# 54395  
DATE: June 13/13  
CUSTOMER: K & S Power Tongs  
MODEL: 6" Tong Torque Gauge  
MANUFACTURER: CDS  
SERIAL NO: # M-7111

SHOULD BE	ACTUAL
Ft.Lbs.	Ft.Lbs.
0	0
7 500	7 500
15 000	15 000
22 500	22 500
30 000	30 000

TECHNICIAN:  QK

COMMENTS: Calibrated for 30 000 ft.lbs. w/ 26" Handle - 4.4301 Comp. Cell

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1

