

ALLWEIGH SCALE SYSTEMS INC
BAY 108, 239 SPRUCE ST.
RED DEER COUNTY, ALBERTA
T4E1B4
(403)309-2220
(403)346-1410
www.allweigh.com



CALIBRATION CERTIFICATE

Customer: **K & S POWER TONGS**
Date: **August 19, 2015**

Model: **Allweigh Digital Torque Indicator**
Serial: **1301019**

Tong Arm Length: **12 Inches**

Calibration Procedure:

- **Load Cell suspended from overhead crane or forklift.**
- **Weights being applied are Government Inspected 1000 Lb Test Weights**
- **Weight is applied to the Tension Load Cell**

Calibration

Weight Applied	Actual Readings	Required Readings
1000 LBS	1000 FT/LBS	1000 FT/LBS
2000 LBS	2000 FT/LBS	2000 FT/LBS
3000 LBS	3000 FT/LBS	3000 FT/LBS

Technician: **Wayne McIlravey**

Calibrations Done at Allweigh Scale Systems Inc Shop