



Certification Of Calibration

CERTIFICATE: 59711
DATE IN SERVICE: June 16, 2016
CUSTOMER: K&S Power Tongs
MODEL: 6" Tong Torque Assembly
MANUFACTURER: Matherne
SERIAL NO. M-19394

SHOULD BE	ACTUAL
Ft. Lbs	Ft. Lbs
0	0
5 000	5 000
10 000	10 000
15 000	15 000
20 000	20 000

Technician Quinton Krause

COMMENTS: 20,000 ft/lbs 24" Handle 2-4.4301 Comp Cells

Test Gauge JW19971106-1 Cal. Date: April 12/16

We hereby certify that on the above date we repaired and re-Calibrated the above instrument and that it is calibrated to Manufacturer's specifications.

The Manufacturer's Specifications State an Accuracy of + - 2% of Full Scale Calibration.
The Test Gauge is Certified to meet and/or Exceed the Requirements as Outlined in API Specifications 6A, Twentieth Edition, October 2010.