

Torque Certification

CERTIFICATE: 53466
DATE: Jan. 17, 2013
CUSTOMER: K&S Power Tongs
MODEL: 6" Tong Torque Gauge
MANUFACTURER: CDS
SERIAL NO: 1041

SHOULD BE Ft.Lbs.	ACTUAL Ft.Lbs.
0	0
3750	3750
7500	7500
11250	11250
15000	15000

TECHNICIAN: T.G. T.G.

COMMENTS: Calibrated for a34" handle & 4.08 tension load cell

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1

