

## Torque Certification

CERTIFICATE: WO# 54494  
DATE: July 6/13  
CUSTOMER: K & S POWER TONGS  
MODEL: 6" Tong Torque Gauge  
MANUFACTURER: CDS  
SERIAL NO: M-13093

SHOULD BE	ACTUAL
Ft. Lbs.	Ft. Lbs.
0	0
2 000	2 000
4 000	4 000
6 000	6 000
8 000	8 000

TECHNICIAN: TC/AVB  
T.G.

COMMENTS: Calibrated for 25" Handle for 4.4301 Compression Cell only.

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1

