



Torque Certification

CERTIFICATE: WO# 54396
DATE: June 13/13
CUSTOMER: K & S Power Tongs
MODEL: 6" Tong Torque Gauge
MANUFACTURER: CDS
SERIAL NO: #M-14270

SHOULD BE	ACTUAL
Ft.Lbs.	Ft.Lbs.
0	0
7 500	7 500
15 000	15 000
22 500	22 500
30 000	30 000

TECHNICIAN:  QK

COMMENTS: Calibrated for 30 000 ft.lbs. w/ 26" Handle - 4.4301 Comp. Cell

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1