

Torque Certification

CERTIFICATE: 58483
DATE: 17-Jun-15
CUSTOMER: K&S Power Tongs
MODEL: 6" Tong Torque Gauge
MANUFACTURER: CDS
SERIAL NO: M16541

| SHOULD BE Ft.Lbs. | ACTUAL Ft.Lbs. |
|----------------------|-------------------|
| 0 | 0 |
| 5000 | 5000 |
| 10000 | 10000 |
| 15000 | 15000 |
| 20000 | 20000 |

TECHNICIAN:  Q.K.

COMMENTS: Calibrated for a 24" handle & 4.4301 sq.in. compression pad
TestGauges/n JW 19971106-1 Cal. Date: 19-May-15

We hereby certify that on the above date we calibrated the above instrument to manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1

