

ALLWEIGH SCALE SYSTEMS INC  
BAY 108, 239 SPRUCE ST.  
RED DEER COUNTY, ALBERTA  
T4E1B4  
(403)309-2220  
(403)346-1410  
www.allweigh.com



## CALIBRATION CERTIFICATE

**Customer:** K & S POWER TONGS  
**Date:** August 19, 2015

**Model:** Allweigh Digital Torque Indicator  
**Serial:** 1211036

**Tong Arm Length:** 12 Inches

**Calibration Procedure:**

- Load Cell suspended from overhead crane or forklift.
- Weights being applied are Government Inspected 1000 Lb Test Weights
- Weight is applied to the Tension Load Cell

**Calibration**

Weight Applied	Actual Readings	Required Readings
1000 LBS	1000 FT/LBS	1000 FT/LBS
2000 LBS	2000 FT/LBS	2000 FT/LBS
3000 LBS	3000 FT/LBS	3000 FT/LBS

**Technician:** Wayne McLravey

Calibrations Done at Allweigh Scale Systems Inc Shop