

ALLWEIGH SCALE SYSTEMS INC
BAY 108, 239 SPRUCE ST.
RED DEER COUNTY, ALBERTA
T4E 1B4
(403)309-2220
(403)346-1410
www.allweigh.com



CALIBRATION CERTIFICATE

Customer: K & S POWER TONGS
Date: August 19, 2015

Model: TC-10K
Serial: 1102013
Calibrated to Computerized Torque Control Units

Calibration Procedure:

- Load Cell suspended from overhead crane or forklift.
- Weights being applied are Government Inspected 1000 Lb Test Weights
- Weight is applied to the Tension Load Cell

Calibration

Weight Applied	Actual Readings
1000 LBS	1000 LBS
2000 LBS	2000 LBS
3000 LBS	3000 LBS
4000 LBS	4000 LBS

Overall calibration accuracy is dependent on line pull angle
System accuracy is +/- 0.5%

Technician: Wayne McIlravey

Calibrations Done at Allweigh Scale Systems Inc Shop