

ALLWEIGH SCALE SYSTEMS INC  
BAY 108, 239 SPRUCE ST.  
RED DEER COUNTY, ALBERTA  
T4E 1B4  
(403)309-2220  
(403)346-1410  
www.allweigh.com



## CALIBRATION CERTIFICATE

**Customer:** K & S POWER TONGS  
**Date:** August 19, 2015

**Model:** TC-10K  
**Serial:** 1303010  
**Calibrated to Computerized Torque Control Units**

### Calibration Procedure:

- Load Cell suspended from overhead crane or forklift.
- Weights being applied are Government Inspected 1000 Lb Test Weights
- Weight is applied to the Tension Load Cell

### Calibration

| Weight Applied | Actual Readings |
|----------------|-----------------|
| 1000 LBS       | 1000 LBS        |
| 2000 LBS       | 2000 LBS        |
| 3000 LBS       | 3000 LBS        |
| 4000 LBS       | 4000 LBS        |

Overall calibration accuracy is dependent on line pull angle  
System accuracy is +/- 0.5%

**Technician:** Wayne McIlravey

Calibrations Done at Allweigh Scale Systems Inc Shop