



Torque Certification

CERTIFICATE: WO# 62068
DATE: December 11th 2017
CUSTOMER: K & S POWER TONGS
MODEL: Tong Torque Gauge
MANUFACTURER: C.D.S
SERIAL NO: # M-6197

SHOULD BE FT.LBS.	ACTUAL FT.LBS.
0	0
7 500	7 500
15 000	15 000
22 500	22 500
30 000	30 000

TECHNICIAN:  Quinton Krause

COMMENTS: Calibrated f/ 0-30 000 FT.LBS. w/ 36" Handle w/ 4.08 Tension Cyl.

Test Gauge JW19971106-1

Cal. Date October 31,2017

We hereby certify that on the above date we calibrated the above instrument to Manufacturer's specifications. Manufacturer's specifications state an accuracy of +/- 2% of full scale calibration. Test gauge is calibrated in compliance with Canadian Standard Association CAN3-Z99-3-85 and A.N.S.I. B40-1